

佛教慈濟醫療財團法人台中慈濟醫院
檢驗醫學科113年量測不確定度審查表

組別：血液組

項目	濃度	量測不確定度(%)	擴充不確定度	comment
WBC	3.04	2.4	4.8	
	7.14	1.7	3.4	
	16.11	1.1	2.2	
RBC	2.34	1.2	2.4	
	4.29	1.0	2	
	5.02	0.8	1.6	
Hb	5.7	1.2	2.4	
	11.7	0.7	1.4	
	15.3	0.7	1.4	
Ht	17.1	1.6	3.2	
	34.6	1.4	2.8	
	44.5	1.2	2.4	
MCV	71.6	1.3	2.6	
	79.9	1.1	2.2	
	87.5	1.2	2.4	
MCH	24.4	1.7	3.4	
	28.4	1.2	2.4	
	30.5	1.3	2.6	
MCHC	33.3	1.8	3.6	
	34.0	1.5	3	
	35.1	1.5	3	
PLT	88	6.6	13.2	
	240	3.1	6.2	
	518	1.6	3.2	
RDW-CV	19.3	1.1	2.2	
	16.5	0.8	1.6	
	15.6	0.8	1.6	
Neutrophil%	41.3	3.0	6	
	43.7	2.3	4.6	
	48.0	2.1	4.2	
Lymphocyte%	34.2	3.8	7.6	
	28.1	3.3	6.6	
	24.8	4.0	8	
Monocyte%	12.2	9.6	19.2	
	9.0	7.5	15	
	11.1	8.1	16.2	
Eosinophil%	9.3	7.8	15.6	
	10.6	7.8	15.6	
	11.7	7.5	15	
Basophil%	4.8	3.9	7.8	
	4.8	3.0	6	
	4.8	2.7	5.4	
NRBC	4.7	10.0	20	
	6.0	5.7	11.4	
	6.6	3.8	7.6	
RET%	5.62	3.5	7	
	2.28	3.5	7	
	1.22	5.0	10	
PT	11.4	2.4	4.8	
	28.2	2.7	5.4	
APTT	28.0	1.9	3.8	
	48.9	1.6	3.2	
	240.0	4.9		
Fibrinogen	79.0	5.7	11.4	
LA Screen	38.4	5.4	10.8	
LA Conform	33.5	6.9	13.8	
D-Dimer	6040.99	8.3	16.6	
	307.63	6.4	12.8	
ESR	8	7.0	14	3SD(來源:AAB)， 以批號2428A換算 之CV%
	17	5.6	11.2	3SD(來源:AAB)， 以批號2428A換算 之CV%
	60	4.0	8	3SD(來源:AAB)， 以批號2428A換算 之CV%
Microbilirubin	11.1	3.3	6.6	
	23.2	3.8	7.6	

E6A0021112-G4

GTA00B002