

佛教慈濟醫療財團法人台中慈濟醫院
檢驗醫學科114年量測不確定度審查表

組別：血液組

項目	濃度	量測不確定度(%)	擴充不確定度	comment
WBC	2.91	2.4	4.82	
	6.66	1.8	3.6	
	16.2	1.4	2.84	
RBC	2.28	1.3	2.64	
	4.21	1.0	1.9	
	5.00	1.0	2	
Hb	5.2	1.9	3.84	
	11.1	0.9	1.8	
	14.8	0.7	1.36	
Ht	15.6	1.9	3.84	
	34.6	1.7	3.46	
	44.1	1.6	3.18	
MCV	71.9	2.1	4.18	
	78.6	2.0	4.08	
	89.1	2.0	4.04	
MCH	23.4	1.3	2.56	
	26.1	1.2	2.3	
	29.5	1.0	2.04	
MCHC	33.6	2.1	4.16	
	34.2	2.1	4.1	
	35.0	2.0	4	
PLT	74	8.1	16.22	
	233	3.0	6	
	514	2.0	3.9	
RDW-CV	18.7	1.1	2.14	
	15.0	0.7	1.34	
	14.8	0.7	1.36	
Neutrophil%	40.5	3.0	5.92	
	45.0	2.4	4.88	
	48.3	2.1	4.14	
Lymphocyte%	32.3	4.6	9.28	
	28.8	3.8	7.64	
	24.8	4.0	8.06	
Monocyte%	12.4	10.5	20.96	
	11.0	10.0	20	
	10.6	9.4	18.86	
Eosinophil%	9.8	8.2	16.32	
	11.1	8.1	16.22	
	11.4	7.9	15.78	
Basophil%	4.7	4.3	8.52	
	4.8	2.1	4.16	
	4.8	2.1	4.16	
NRBC	4.5	11.1	22.22	
	6.2	6.5	12.9	
	6.4	3.1	6.26	
RET%	5.92	4.6	9.12	
	2.41	5.4	10.78	
	1.24	7.3	14.52	
PT	11.1	1.8	3.6	
	28.2	2.8	5.68	
APTT	28.5	2.5	4.92	
	47.0	2.3	4.68	
Fibrinogen	241.6	5.6	11.1	
	67.7	6.5	13	
LA Screen	34.6	4.3	8.68	
LA Conform	34.6	4.6	9.24	
D-Dimer	5233.12	4.8	9.68	
	257.06	5.1	10.22	
ESR	9	12.5	25	3SD(來源:AAB), 以批號2428A換算 之CV%
	18	5.9	11.76	3SD(來源:AAB), 以批號2428A換算 之CV%
	62	3.3	6.66	3SD(來源:AAB), 以批號2428A換算 之CV%
Microbilirubin	10.8	2.8	5.5	
	25.1	3.2	6.32	

E6A0021112-G4

GTA00B002

已於 115/03/26 主管會議經主管確認無誤