

佛教慈濟醫療財團法人台中慈濟醫院
檢驗醫學科量測不確定度審查表

年度:2025 組別:血液組

項目	濃度	量測不確定度(1TEa(%))	1/3TEa(%)	目標不確定度 (goal)	允收判定	審查	comment
WBC	3.04	2.4	15	5.0	<1/3 TEa	5.0	●合格
	7.14	1.7	15	5.0	<1/3 TEa	5.0	●合格
	16.11	1.1	15	5.0	<1/3 TEa	5.0	●合格
RBC	2.34	1.2	6	2.0	<1/3 TEa	2.0	●合格
	4.29	1.0	6	2.0	<1/3 TEa	2.0	●合格
	5.02	0.8	6	2.0	<1/3 TEa	2.0	●合格
Hb	5.7	1.2	7	2.3	<1/3 TEa	2.3	●合格
	11.7	0.7	7	2.3	<1/3 TEa	2.3	●合格
	15.3	0.7	7	2.3	<1/3 TEa	2.3	●合格
Ht	17.1	1.6	6	2.0	<1/3 TEa	2.0	●合格
	34.6	1.4	6	2.0	<1/3 TEa	2.0	●合格
	44.5	1.2	6	2.0	<1/3 TEa	2.0	●合格
MCV	71.6	1.3	8	2.7	<1/3 TEa	2.7	●合格
	79.9	1.1	8	2.7	<1/3 TEa	2.7	●合格
	87.5	1.2	8	2.7	<1/3 TEa	2.7	●合格
MCH	24.4	1.7	9	3.0	<1/3 TEa	3.0	●合格
	28.4	1.2	9	3.0	<1/3 TEa	3.0	●合格
	30.5	1.3	9	3.0	<1/3 TEa	3.0	●合格
MCHC	33.3	1.8	9	3.0	<1/3 TEa	3.0	●合格
	34.0	1.5	9	3.0	<1/3 TEa	3.0	●合格
	35.1	1.5	9	3.0	<1/3 TEa	3.0	●合格
PLT	88	6.6	25	8.3	<1/3 TEa	8.3	●合格
	240	3.1	25	8.3	<1/3 TEa	8.3	●合格
	518	1.6	25	8.3	<1/3 TEa	8.3	●合格
RDW-CV	19.3	1.1	4.6	1.5	<1/3 TEa	1.5	●合格
	16.5	0.8	4.6	1.5	<1/3 TEa	1.5	●合格
	15.6	0.8	4.6	1.5	<1/3 TEa	1.5	●合格
Neutrophil%	41.3	3.0	22.4	7.5	<1/3 TEa	7.5	●合格
	43.7	2.3	22.4	7.5	<1/3 TEa	7.5	●合格
	48.0	2.1	22.4	7.5	<1/3 TEa	7.5	●合格
Lymphocyte%	34.2	3.8	16.0	5.3	<1/3 TEa	5.3	●合格
	28.1	3.3	16.0	5.3	<1/3 TEa	5.3	●合格
	24.8	4.0	16.0	5.3	<1/3 TEa	5.3	●合格
Monocyte%	12.2	9.6	40	13.3	<1/3 TEa	13.3	●合格
	9.0	7.5	40	13.3	<1/3 TEa	13.3	●合格
	11.1	8.1	40	13.3	<1/3 TEa	13.3	●合格
Eosinophil%	9.3	7.8	37.1	12.4	<1/3 TEa	12.4	●合格
	10.6	7.8	37.1	12.4	<1/3 TEa	12.4	●合格
	11.7	7.5	37.1	12.4	<1/3 TEa	12.4	●合格
Basophil%	4.8	3.9	38.5	12.8	<1/3 TEa	12.8	●合格
	4.8	3.0	38.5	12.8	<1/3 TEa	12.8	●合格
	4.8	2.7	38.5	12.8	<1/3 TEa	12.8	●合格
NRBC	4.7	10.0	50	16.7	<1/3 TEa	16.7	●合格
	6.0	5.7	50	16.7	<1/3 TEa	16.7	●合格
	6.6	3.8	50	16.7	<1/3 TEa	16.7	●合格
RET%	5.62	3.5	30	10.0	<1/3 TEa	10.0	●合格
	2.28	3.5	30	10.0	<1/3 TEa	10.0	●合格
	1.22	5.0	30	10.0	<1/3 TEa	10.0	●合格
PT	11.4	2.4	15	5.0	<1/3 TEa	5.0	●合格
	28.2	2.7	15	5.0	<1/3 TEa	5.0	●合格
	28.0	1.9	15	5.0	<1/3 TEa	5.0	●合格
APTT	48.9	1.6	15	5.0	<1/3 TEa	5.0	●合格
	240.0	4.9	20	6.7	<1/3 TEa	6.7	●合格
	79.0	5.7	20	6.7	<1/3 TEa	6.7	●合格
LA Screen	38.4	5.4	30	10.0	<1/3 TEa	10.0	●合格
LA Conform	33.5	6.9	30	10.0	<1/3 TEa	10.0	●合格
D-Dimer	6040.99	8.3	30	10.0	<1/3 TEa	10.0	●合格
	307.63	6.4	30	10.0	<1/3 TEa	10.0	●合格
ESR	8	7.0	33	16.5	<1/2 TEa	16.5	●合格 3SD(來源:AAB), 以批號2428A換算 之CV%
	17	5.6	15	7.5	<1/2 TEa	7.5	●合格 3SD(來源:AAB), 以批號2428A換算 之CV%
	60	4.0	9	4.5	<1/2 TEa	4.5	●合格 3SD(來源:AAB), 以批號2428A換算 之CV%
Microbilirubin	11.1	3.3	20	6.7	<1/3 TEa	6.7	●合格
	23.2	3.8	20	6.7	<1/3 TEa	6.7	●合格