

佛教慈濟醫療財團法人台中慈濟醫院

檢驗醫學科量測不確定度審查表

年度：114 / 組別：血免疫清組

項目	Level	單位	濃度	量測不確定度(%)	目標不確定度 (goal)	允收判定	符合判定
Tacrolimus	Level 1	ng/mL	5.11	5.89	<2CV	11.8	<2CV
	Level 2	ng/mL	9.19	6.30	<2CV	12.6	<2CV
	Level 3	ng/mL	19.96	7.65	<2CV	15.3	<2CV
Cyclosporine	Level 1	ng/mL	95.90	8.05	<2CV	16.1	<2CV
	Level 2	ng/mL	347.21	6.66	<2CV	13.3	<2CV
	Level 3	ng/mL	806.17	9.22	<2CV	18.4	<2CV
ACTH	Level 1	pg/mL	41.44	6.05	<2CV	12.1	<2CV
	Level 2	pg/mL	242.34	5.85	<2CV	11.7	<2CV
Aldo	Level 1	ng/dL	6.23	10.10	<2CV	20.2	<2CV
	Level 2	ng/dL	28.80	7.66	<2CV	15.3	<2CV
GH	Level 1	ng/mL	3.79	4.02	<2CV	8.0	<2CV
	Level 2	ng/mL	18.71	3.83	<2CV	7.7	<2CV
DHEA-S	Level 1	µg/dL	45.2	7.74	<2CV	15.5	<2CV
	Level 2	µg/dL	157.6	6.66	<2CV	13.3	<2CV
T-IgE	Level 1	KU/L	21.29	6.44	<2CV	12.9	<2CV
	Level 2	KU/L	86.15	5.79	<2CV	11.6	<2CV
	Level 3	KU/L	411.91	5.60	<2CV	11.2	<2CV
HBsAg定量	Level 1	IU/ml	0.00	0.00	<2CV	0.0	<2CV
	Level 2	IU/ml	0.19	4.39	<2CV	8.8	<2CV
	Level 3	IU/ml	28.37	4.58	<2CV	9.2	<2CV
PIVKA-II	Level 1	mAU/ml	43.50	7.76	<2CV	15.5	<2CV
	Level 2	mAU/ml	188.60	8.12	<2CV	16.2	<2CV
Anti-HIV	Level 1	S/CO	4.77	7.47	<2CV	14.9	<2CV
	Level 2	S/CO	3.77	9.08	<2CV	18.2	<2CV
	Level 3	S/CO	3.15	3.71	<2CV	7.4	<2CV
Mycoplasma IgM	Level 1	Index	0.00	0.00	<2CV	0.0	<2CV
	Level 2	Index	20.97	12.83	<2CV	25.7	<2CV
VZV-G	Level 1	mIU/mL	46.84	11.26	<2CV	22.5	<2CV
	Level 2	mIU/mL	451.19	10.56	<2CV	21.1	<2CV
Measles IgG	Level 1	mIU/mL	0.00	0.00	<2CV	0.0	<2CV
	Level 2	mIU/mL	84.35	11.63	<2CV	23.3	<2CV
HSV-1/2 IgM	Level 1	Index	0.00	0.00	<2CV	0.0	<2CV
	Level 2	Index	2.40	22.19	<2CV	44.4	<2CV
HSV-1/2 IgG	Level 1	Index	0.00	0.00	<2CV	0.0	<2CV
	Level 2	Index	7.65	13.82	<2CV	27.6	<2CV
Mumps Virus IgG	Level 1	mIU/mL	0.00	0.00	<2CV	0.0	<2CV
	Level 2	mIU/mL	64.30	9.10	<2CV	18.2	<2CV
Anti-Cardiolipin IgG	NEG	GPL-U/m	0.00	0.00	<2CV	0.0	<2CV
	POS	GPL-U/m	40.08	3.80	<2CV	7.6	<2CV
Anti-cardiolipin-IgM	NEG	MPL-U/m	0.00	0.00	<2CV	0.0	<2CV
	POS	MPL-U/m	61.95	4.11	<2CV	8.2	<2CV
Anti-β2-glycoprotein-I	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	54.52	3.76	<2CV	7.5	<2CV
Anti-β2-glycoprotein-I	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	52.08	5.02	<2CV	10.0	<2CV
Anti-ENA Scl-70 Ab	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	33.64	7.43	<2CV	14.9	<2CV
Anti-ENA Smd	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	42.08	5.28	<2CV	10.6	<2CV
Anti-ENA RNP	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	60.29	5.48	<2CV	11.0	<2CV
Anti-ENA SS-A(Ro)	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	64.54	6.22	<2CV	12.4	<2CV
Anti-ENA SS-B(La)	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	40.21	4.74	<2CV	9.5	<2CV
RF-IgA	NEG	IU/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	IU/ml	59.67	5.78	<2CV	11.6	<2CV
RF-IgM	NEG	IU/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	IU/ml	36.34	6.63	<2CV	13.3	<2CV
Anti-ds DNA Antibody	NEG	IU/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	IU/ml	47.32	9.23	<2CV	18.5	<2CV
Anti Jo-1 antibody	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	36.22	4.95	<2CV	9.9	<2CV
Anti CENPB	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	36.15	3.75	<2CV	7.5	<2CV
Anti-CTD	NEG	ratio	0.00	0.00	<2CV	0.0	<2CV
	POS	ratio	10.80	7.60	<2CV	15.2	<2CV
Anti-CCP	NEG	U/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	U/ml	46.19	4.98	<2CV	10.0	<2CV
PR3	NEG	IU/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	IU/ml	11.81	9.34	<2CV	18.7	<2CV
MPO	NEG	IU/ml	0.00	0.00	<2CV	0.0	<2CV
	POS	IU/ml	23.52	4.99	<2CV	10.0	<2CV